




VIBRATION ANALYSIS HARDWARE



**"V" Series MIL-Style Boot Connector Kits  
Product Manual**


## TABLE OF CONTENTS

- Introduction.....3
- Assembly for V2X Series and CK-JV2X Series.....6
- Assembly for V3X Series.....9
- Assembly for V4X Series and CK-JV4X Series..... 12
- Maintenance..... 15
- Warranty & Return Information..... 15




# INTRODUCTION


This document contains information on the operation, installation and maintenance of the V-style series of connector kits.









**Figure 1. 2-Socket Connector Kit Materials**




**Figure 2. 3-Socket Connector Kit Materials**



**Figure 2. 4-Socket Connector Kit Materials**

Part #	Connector Kit Parts	Material (Insert)	Max Temp.	Connector Kit Parts #	Compatible Cables
JV2R	 CK-JV2R	Polyphenylene Sulfide, VITON® Boot	350 °F (177 °C)	CK-JV2R	CB111
V2J	 CK-V2J	Polycarbonate, VITON® Boot	250 °F (121 °C)	CK-V2J	CB111
V2R	 CK-V2R	Polyphenylene Sulfide, VITON® Boot	350 °F (177 °C)	CK-V2R	CB111
V3R	 CK-V3R	Polyphenylene Sulfide, VITON® Boot	350 °F (177 °C)	CK-V3R	CB112
JV4R	 CK-JV4R	Polyphenylene Sulfide, VITON® Boot	350 °F (177 °C)	CK-JV4R	CB119
V4J	 CK-V4J	Polycarbonate, VITON® Boot	250 °F (121 °C)	CK-V4J	CB119

**Table 1. Product Selection Guide**



Recommended Tool: CB926-1A




Recommended Epoxy: MH109-3D



## ASSEMBLY FOR CK-V2X SERIES AND CK-JV2X SERIES


1. Strip outer jacket of wire .300 in.



2. Cut off shield and drain wire (for twisted shielded wires only).



3. Strip the insulation of two conductor wires back .200 in.




4. Crimp conductor wires into contact sockets. CTC's CB926-1A crimp tool makes crimping fast and easy and can lead to significant time savings when installing a large volume of connector kits. Adjusting the green depth knob to the desired length allows the depth of the contacts to be set manually to ensure a crimp at the correct location every time. Suggested depth for the V2J is 0.43 in. Suggested depth for the V2R is 0.52 in.




5. Slide VITON® boot and backshell onto the cable.




6. Using a fine-tipped punch, gently press each contact into the appropriate position on the insert.
  - a. Install accelerometer red (+) wire into the insert socket for Pin A.
  - b. Install accelerometer black (-) wire into insert socket for Pin B.



7. Thread the backshell onto the insert.



8. Place the assembled connector body horizontally with the two small epoxy injection holes level and facing upward.
9. Mix epoxy. Using a syringe, fill the backshell with epoxy through one of the small injection holes until epoxy begins to seep from the other.



10. Keep the connector in a horizontal position, allowing the epoxy to set and vent any trapped air, refilling as needed.

11. Allow the epoxy to cure for six hours at room temperature. Place a piece of masking tape over the two epoxy holes to prevent leakage and hang the connector vertically with the insert facing downward. This will ensure epoxy encapsulates the cable evenly, especially if the cable diameter is smaller than the backshell opening.



12. Push the VITON® boot shell over the assembled connector body until it snaps into place.




The connector body should be recessed approximately 0.125 in. from the tip of the VITON® boot when properly installed.






## ASSEMBLY FOR CK-V3X SERIES


1. Strip outer jacket of wire .300 in.



2. Cut off shield and drain wire (for twisted shielded wires only).



3. Strip the insulation of three conductor wires back .200 in.




4. Crimp conductor wires into contact sockets. CTC's CB926-1A crimp tool makes crimping fast and easy and can lead to significant time savings when installing a large volume of connector kits. Adjusting the green depth knob to the desired length allows the depth of the contacts to be set manually to ensure a crimp at the correct location every time. Suggested depth for the V3R is 0.52 in.




5. Slide VITON® boot and backshell onto the cable.




6. Attach O-ring to the front of the insert.
7. Using a fine-tipped punch, gently press each contact into the appropriate position on the insert.
  - a. Install accelerometer red (+) wire into the insert socket for Pin A.
  - b. Install accelerometer black (-) wire into insert socket for Pin B.
  - c. Install accelerometer white (+) wire into the insert socket for Pin C.



8. Thread the backshell onto the insert.



9. Place the assembled connector body horizontally with the two small epoxy injection holes level and facing upward.
10. Mix epoxy. Using a syringe, fill the backshell with epoxy through one of the small injection holes until epoxy begins to seep from the other.



11. Keep the connector in a horizontal position, allowing the epoxy to set and vent any trapped air, refilling as needed.

12. Allow the epoxy to cure for six hours at room temperature. Place a piece of masking tape over the two epoxy holes to prevent leakage and hang the connector vertically with the insert facing downward. This will ensure epoxy encapsulates the cable evenly, especially if the cable diameter is smaller than the backshell opening.



13. Push the VITON® boot shell over the assembled connector body until it snaps into place.




The connector body should be recessed approximately 0.125 in. from the tip of the VITON® boot when properly installed.



## ASSEMBLY FOR CK-V4X SERIES AND CK-JV4X SERIES


1. Strip outer jacket of wire .300 in.



2. Cut off shield and drain wire (for twisted shielded wires only).



3. Strip the insulation of two conductor wires back .200 in.




4. Crimp conductor wires into contact sockets. CTC's CB926-1A crimp tool makes crimping fast and easy and can lead to significant time savings when installing a large volume of connector kits. Adjusting the green depth knob to the desired length allows the depth of the contacts to be set manually to ensure a crimp at the correct location every time. Suggested depth for the V4J is 0.43 in.




5. Slide VITON® boot and backshell onto the cable.




6. Using a fine-tipped punch, gently press each contact into the appropriate position on the insert.
  - a. Install accelerometer red (+) wire into the insert socket for Pin A.
  - b. Install accelerometer green (+) wire into the insert socket for Pin B.
  - c. Install accelerometer white (+) wire into the insert socket for Pin C.
  - d. Install accelerometer black (-) wire into insert socket for Pin D.



7. Thread the backshell onto the insert.



8. Place the assembled connector body horizontally with the two small epoxy injection holes level and facing upward.
9. Mix epoxy. Using a syringe, fill the backshell with epoxy through one of the small injection holes until epoxy begins to seep from the other.



10. Keep the connector in a horizontal position, allowing the epoxy to set and vent any trapped air, refilling as needed.
11. Allow the epoxy to cure for six hours at room temperature. Place a piece of masking tape over the two epoxy holes to prevent leakage and hang the connector vertically with the insert facing downward. This will ensure epoxy encapsulates the cable evenly, especially if the cable diameter is smaller than the backshell opening.



12. Push the VITON® boot shell over the assembled connector body until it snaps into place.



The connector body should be recessed approximately 0.125 in. from the tip of the VITON® boot when properly installed.



## **MAINTENANCE**

Once the product has been installed, minimal maintenance will be required. Basic visual checks to ensure integrity should be made periodically.

### **General**

There are no customer-replaceable parts. The product has been designed for trouble-free service under normal operating conditions.

## **WARRANTY & REFUND**

Please visit [www.ctconline.com](http://www.ctconline.com) to view a complete recapitulation of our warranty and refund policies.

## **CONTACT INFORMATION**

Connection Technology Center, Inc. (CTC)

7939 Rae Blvd.

Victor, NY 14564

1.585.924.5900

[sales@ctconline.com](mailto:sales@ctconline.com)

[www.ctconline.com](http://www.ctconline.com)

